



# LATEX

## Examination Gloves

### INFORMATION

- High grade quality latex compound
- Medical grade AQL 1.5 compliance
- Protection from unwanted or dangerous substances
- Available in absorbable USP grade powder, chlorinated and polymer coated
- Excellent fit, easy donning
- Biocompatibility tested

### PHYSICAL DIMENSION AND PROPERTIES :

PHYSICAL DIMENSION	ASTM D3578	EN 455-2
<b>Length (mm)</b>		
	(XS, S) Min. 220 (M, L, XL) Min. 230	Min. 240
<b>Thickness (mm)</b>		
Finger	Min. 0.08	N/A
Palm	Min. 0.08	N/A
<b>Palm Width (mm)</b>		
XS	70 ± 10	≤ 80
S	80 ± 10	80 ± 10
M	95 ± 10	95 ± 10
L	111 ± 10	110 ± 10
XL	120 ± 10	≥ 110

PHYSICAL PROPERTIES	ASTM D3578	EN 455-2
<b>Tensile Strength (MPa)</b>		
Before Aging	Min. 18 MPa	N/A
After Aging	Min. 14 MPa	N/A
<b>Elongation at Break (%)</b>		
Before Aging	Min. 650	N/A
After Aging	Min. 500	N/A
<b>Force at Break (N)</b>		
Before Aging	N/A	Min. 6
After Aging	N/A	Min. 6




### Quality Standard:

- ISO 13485- Quality Management System - Medical Device
- In compliance with European Medical Device Directive 93/42/EEC (CE Class I)
- In compliance with Personal Protective Equipment, PPE Regulation (EU) 2016/425
- ASTM D3578

### Product Specification

- **Type:**  
Powdered or Powder Free
- **Sterilization:**  
Non-Sterile
- **Colour:**  
Natural
- **Design**  
Ambidextrous, textured/smooth surface, and beaded cuff

 **Bergamot Sdn Bhd** (1153028-V)  
Taman Industri Balakong Jaya,  
43300 Seri Kembangan,  
Selangor, Malaysia

 [info@bergamot.com.my](mailto:info@bergamot.com.my)

 +603 8964 4186

