



# LATEX Z SERIES Examination Gloves

## INFORMATION

- High grade quality latex compound
- Medical grade AQL 1.5 compliance
- Protection from unwanted or dangerous substances
- Available in absorbable USP grade powder, chlorinated and polymer coated
- Excellent fit, easy donning
- Biocompatibility tested

### PHYSICAL DIMENSION AND PROPERTIES :

PHYSICAL DIMENSION	ASTM D3578	EN 455-2
	<b>Length (mm)</b>	
	(XS, S) Min. 220	Min. 240
	(M, L, XL) Min. 230	
	<b>Thickness (mm)</b>	
Finger	Meeting ASTM & EN Standards.	
Palm	Request for further product information.	
	<b>Palm Width (mm)</b>	
XS	70 ± 10	≤ 80
S	80 ± 10	80 ± 10
M	95 ± 10	95 ± 10
L	111 ± 10	110 ± 10
XL	120 ± 10	≥ 110

PHYSICAL PROPERTIES	ASTM D3578	EN 455-2
	<b>Tensile Strength (MPa)</b>	
Before Aging	Min. 18 MPa	N/A
After Aging	Min. 14 MPa	N/A
	<b>Elongation at Break (%)</b>	
Before Aging	Min. 650	N/A
After Aging	Min. 500	N/A
	<b>Force at Break (N)</b>	
Before Aging	N/A	Min. 6
After Aging	N/A	Min. 6




### Quality Standard:

- ISO 13485- Quality Management System - Medical Device
- In compliance with Medical Device Regulation (EU), MDR 2017/745 (Class 1)
- In compliance with Personal Protective Equipment, PPE Regulation (EU) 2016/425, Cat III
- ASTM D3578

### Product Specification

- **Type:**  
Powdered or Powder Free
- **Sterilization:**  
Non-Sterile
- **Colour:**  
Natural
- **Design**  
Ambidextrous, textured/smooth surface, and beaded cuff

 **Bergamot Sdn Bhd** (1153028-V)  
No 7 & 8, Jalan Sinaran 6,  
Kawasan Perindustrian Sinaran,  
43000 Kajang, Selangor, Malaysia.

 [info@bergamot.com.my](mailto:info@bergamot.com.my)

 +603 8964 4186