



NITRILE X SERIES

Examination Gloves

INFORMATION

- Formulated with Acrylonitrile-butadiene, alternative solution for Type 1 allergic individuals
- Medical grade AQL 1.5 compliance
- Protection from unwanted or dangerous substances
- Biocompatibility tested
- Enhanced soft comfort with excellent grip
- Tactile sensitivity

PHYSICAL DIMENSION AND PROPERTIES :

PHYSICAL DIMENSION	ASTM D6319	EN 455-2
	Length (mm)	
	(XS, S) Min. 220	Min. 240
	(M, L, XL) Min. 230	
	Thickness (mm)	
Finger	Meeting ASTM & EN Standards.	
Palm	Request for further product information.	
	Palm Width (mm)	
XS	70 ± 10	≤ 80
S	80 ± 10	80 ± 10
M	95 ± 10	95 ± 10
L	110 ± 10	110 ± 10
XL	120 ± 10	≥ 110

PHYSICAL PROPERTIES	ASTM D6319	EN 455-2
	Tensile Strength (MPa)	
Before Aging	Min. 14 MPa	N/A
After Aging	Min. 14 MPa	N/A
	Elongation at Break (%)	
Before Aging	Min. 500	N/A
After Aging	Min. 400	N/A
	Force at Break (N)	
Before Aging	N/A	Min. 6
After Aging	N/A	Min. 6



COLOUR AVAILABLE :



Quality Standard:

- ISO 13485- Quality Management System - Medical Device
- In compliance with Medical Device Regulation (EU), MDR 2017/745 (Class 1)
- In compliance with Personal Protective Equipment, PPE Regulation (EU) 2016/425, Cat III
- ASTM D6319

Product Specification

- **Type:**
Powder Free
- **Sterilization:**
Non-Sterile
- **Colour:**
Blue, Black, Violet, Purple, Pink, and White
- **Design**
Ambidextrous, textured, and beaded cuff

Bergamot Sdn Bhd (1153028-V)
No 7 & 8, Jalan Sinaran 6,
Kawasan Perindustrian Sinaran,
43000 Kajang, Selangor, Malaysia.

info@bergamot.com.my

+603 8964 4186